
# Field Evaluation Sensirion SEN54



## Background

 From 11/08/2023 to 12/10/2023, and then from 6/14/2024 to 7/19/2024, three Sensirion SEN54 units were deployed at the South Coast AQMD stationary ambient monitoring site in Rubidoux and were run side-by-side with Federal Equivalent Method (FEM) instruments measuring the same pollutants.

#### <u>Sensirion SEN54 (3 units tested)</u>:

- PM Optical (Sensirion SEN54, non-FEM)
- Each unit measures: PM<sub>1.0</sub> (µg/m<sup>3</sup>), PM<sub>2.5</sub> (µg/m<sup>3</sup>), PM<sub>10</sub>(µg/m<sup>3</sup>), T (°C), RH (%)
- > Also reports:  $PM_{4.0}$  (µg/m<sup>3</sup>), VOC index
- ➤ Unit cost: ~\$23
- Time resolution: 1 second
- Units IDs: 2B7D, 1F07, and 415B





**Sensirion SEN54** 


- South Coast AQMD Reference instruments:
- Teledyne API T640 (hereinafter FEM T640 for PM<sub>2.5</sub>, T640 otherwise):
  - > Optical particle counter (FEM PM<sub>2.5</sub>)
  - > Measures  $PM_{1.0}$ ,  $PM_{2.5}$  and  $PM_{10}$  (µg/m<sup>3</sup>)
  - ≻ Cost: ~\$21,000
  - ➤ Time resolution: 1-min
- MetOne BAM:
  - ➢ Beta-attenuation (FEM PM<sub>2.5</sub> & PM<sub>10</sub>)
  - > Measures  $PM_{2.5}$ , and  $PM_{10}$  (µg/m<sup>3</sup>)
  - ≻ Cost: ~\$20,000
  - ➤ Time resolution: 1-hr
- Met Station (T, RH, P, WS, WD):
  - ≻ Cost: ~\$5,000
  - ➤ Time resolution: 1-min

### Data validation & recovery

- Basic QA/QC procedures were used to validate the collected data (i.e. obvious outliers, negative values and invalid data-points were eliminated from the data-set)
- Data recovery from Unit 2B7D, Unit 1F07 and Unit 415B was ~94.5% respectively for all PM measurements
- Data related to 4<sup>th</sup> of July activities were excluded from data analysis for all sensors and reference instruments


### Sensirion SEN54; intra-model variability

- Absolute intra-model variability was ~0.11, ~0.06 and ~0.05 µg/m<sup>3</sup> for PM<sub>1.0</sub>, PM<sub>2.5</sub> and PM<sub>10</sub>, respectively (calculated as the standard deviation of the three sensor means)
- Relative intra-model variability was ~1.09%, ~0.56% and ~0.46% for PM<sub>1.0</sub>, PM<sub>2.5</sub> and PM<sub>10</sub>, respectively (calculated as the absolute intra-model variability relative to the mean of the three sensor means)




#### Reference Instruments: PM<sub>2.5</sub> FEM BAM and FEM T640

- Data recovery for PM<sub>2.5</sub> from FEM BAM and FEM T640 was ~ 91.1% and 99.8%, respectively.
- Strong correlations between the reference instruments for PM<sub>2.5</sub> measurements (R<sup>2</sup> ~0.75) were observed.




#### Reference Instruments: PM<sub>10</sub> FEM BAM and T640


- Data recovery for  $PM_{10}$  from FEM BAM and T640 was ~ 93.9% and 99.7%, respectively.
- Strong correlations between the reference instruments for  $PM_{10}$  measurements (R<sup>2</sup> ~0.85) were observed.




#### Sensirion SEN54 vs T640 (PM<sub>1.0</sub>; 5-min mean)




- The Sensirion SEN54 sensors showed strong correlations with the corresponding T640 data (0.86 < R<sup>2</sup> < 0.89)</li>
- Overall, the Sensirion SEN54 sensors overestimated the PM<sub>1.0</sub> mass concentrations as measured by T640
- The Sensirion SEN54 sensors seemed to track the PM<sub>1.0</sub> diurnal variations as recorded by T640




#### Sensirion SEN54 vs FEM T640 (PM<sub>2.5</sub>; 5-min mean)




- The Sensirion SEN54 sensors showed strong correlations with the corresponding FEM T640 data (0.76 < R<sup>2</sup> < 0.77)</li>
- Overall, the Sensirion SEN54 sensors underestimated the PM<sub>2.5</sub> mass concentrations as measured by FEM T640
- The Sensirion SEN54 sensors seemed to track the PM<sub>2.5</sub> diurnal variations as recorded by FEM T640




#### Sensirion SEN54 vs T640 (PM<sub>10</sub>; 5-min mean)




- The Sensirion SEN54 sensors showed very weak correlations with the corresponding T640 data (0.18 < R<sup>2</sup> < 0.20)</li>
- Overall, the Sensirion SEN54 sensors underestimated the PM<sub>10</sub> mass concentrations as measured by T640
- The Sensirion SEN54 sensors did not seem to track the PM<sub>10</sub> diurnal variations as recorded by T640




#### Sensirion SEN54 vs T640 (PM<sub>1.0</sub>; 1-hr mean)




#### Sensirion SEN54 vs FEM T640 (PM<sub>2.5</sub>; 1-hr mean)




#### Sensirion SEN54 vs T640 (PM<sub>10</sub>; 1-hr mean)




- The Sensirion SEN54 sensors showed very weak correlations with the corresponding T640 data (0.20 < R<sup>2</sup> < 0.22)</li>
- Overall, the Sensirion SEN54 sensors underestimated the PM<sub>10</sub> mass concentrations as measured by T640
- The Sensirion SEN54 sensors did not seem to track the PM<sub>10</sub> diurnal variations as recorded by T640



#### Sensirion SEN54 vs T640 (PM<sub>1.0</sub>; 24-hr mean)




#### Sensirion SEN54 vs FEM T640 (PM<sub>2.5</sub>; 24-hr mean)




<sup>13</sup> 

#### Sensirion SEN54 vs T640 (PM<sub>10</sub>; 24-hr mean)




14


#### Sensirion SEN54 vs FEM BAM (PM<sub>2.5</sub>; 1-hr mean)




- The Sensirion SEN54 sensors showed moderate correlations with the corresponding FEM BAM data (0.51 < R<sup>2</sup> < 0.53)</li>
- Overall, the Sensirion SEN54 sensors underestimated the PM<sub>2.5</sub> mass concentrations as measured by FEM BAM
- The Sensirion SEN54 sensors seemed to track the PM<sub>2.5</sub> diurnal variations as recorded by FEM BAM




#### Sensirion SEN54 vs FEM BAM (PM<sub>10</sub>; 1-hr mean)




- The Sensirion SEN54 sensors showed no to very weak correlations with the corresponding FEM BAM data (0.09 < R<sup>2</sup> < 0.11)</li>
- Overall, the Sensirion SEN54 sensors underestimated the PM<sub>10</sub> mass concentrations as measured by FEM BAM
- The Sensirion SEN54 sensors did not seem to track the PM<sub>10</sub> diurnal variations as recorded by FEM BAM



#### Sensirion SEN54 vs FEM BAM (PM<sub>2.5</sub>; 24-hr mean)



#### Sensirion SEN54 vs FEM BAM (PM<sub>10</sub>; 24-hr mean)



## Summary: PM

	Average of 3 Sensors, PM <sub>1.0</sub>		Sensirion SEN54 vs T640, PM <sub>1.0</sub>						T640 (PM <sub>1.0</sub> , μg/m <sup>3</sup> )		
	Average (µg/m³)	SD (µg/m³)	R <sup>2</sup>	Slope	Intercept	MBE <sup>1</sup> (µg/m <sup>3</sup> )	MAE <sup>2</sup> (µg/m <sup>3</sup> )	RMSE <sup>3</sup> (µg/m <sup>3</sup> )	Ref. Averag	e Ref. SD	Range during the field evaluation
5-min	10.1	7.8	0.87 to 0.88	0.62 to 0.66	2.6 to 2.8	0.8 to 1.0	2.1 to 2.4	3.2 to 3.8	9.0	5.4	0.8 to 38.2
1-hr	10.1	7.7	0.88 to 0.90	0.62 to 0.67	2.5 to 2.8	0.8 to 1.0	2 to 2.3	3.1 to 3.6	9.0	5.3	0.9 to 33.9
24-hr	10.2	5.9	0.93	0.65 to 0.67	2.5 to 2.6	0.8 to 1.1	1.7 to 1.8	2.4 to 2.5	9.0	4.1	1.2 to 19.8
	Average of 3 Sensors, PM <sub>2.5</sub>		Sensirion SEN54 vs FEM BAM & FEM T640, PM <sub>2.5</sub>						FEM BAM & FEM T640 (PM <sub>2.5</sub> , μg/m <sup>3</sup> )		
	Average (µg/m³)	SD (µg/m³)	R <sup>2</sup>	Slope	Intercept	MBE <sup>1</sup> (µg/m <sup>3</sup> )	MAE <sup>ź</sup> (µg/m		Ref. Average	Ref. SD	Range during the field evaluation
5-min	10.8	8.3	0.76 to 0.77	0.67 to 0.69	5.0 to 5.2	-1.7 to -1.6	3.52 to 3	.54 4.4 to 4.5	12.2	6.5	1.1 to 50.4
1-hr	10.7	8.2	0.52 to 0.78	0.55 to 0.69	5.0 to 5.6	-1.7 to -0.6	3.5 to 4	.2 4.3 to 5.8	11.6 to 12.2	6.3 to 6.4	0 to 41.5
24-hr	10.9	6.3	0.59 to 0.78	0.55 to 0.68	5.0 to 5.7	-1.7 to -0.7	2.9 to 3	.4 3.4 to 4.0	11.5 to 12.2	4.5 to 4.9	2.6 to 24.4
	Averaç Sensor	•	Sensirion SEN54 vs FEM BAM & T640, PM <sub>10</sub>						FEM BAM & T640 (PM <sub>10</sub> , μg/m <sup>3</sup> )		
	Average (µg/m³)	SD (µg/m³)	R <sup>2</sup>	Slope	Intercept	MBE <sup>1</sup> (µg/m <sup>3</sup> )	MAE <sup>2</sup> (µg/m <sup>3</sup> )	RMSE <sup>3</sup> (µg/m <sup>3</sup> )	Ref. Average	e Ref. SE	Range during the field evaluation
5-min	11	8.5	0.18 to 0.19	0.93 to 0.97	29.0 to 29.3	-28.6 to -28.5	28.6 to 28.7	33.0 to 33.2	39.7	19.0	2.1 to 366.9
1-hr	11.0	8.4	0.10 to 0.22	0.69 to 0.96	29.0 to 30.2	-28.6 to -26.7	27.3 to 28.7	32.2 to 32.5	38.2 to 39.7	17.7 to 19	0.5 1.9 to 206.7
24-hr	11.1	6.4	0.18 to 0.25	0.76 to 0.92	28.3 to 30.3	-28.5 to -26.3	26.4 to 28.5	28.3 to 30.2	38.1 to 39.7	11.4 to 12	2.0 12.0 to 65.4


<sup>1</sup>Mean Bias Error (MBE): the difference between the sensors and the reference instruments. MBE indicates the tendency of the sensors to underestimate (negative MBE values) or overestimate (positive MBE values).

<sup>2</sup> Mean Absolute Error (MAE): the absolute difference between the sensors and the reference instruments. The larger MAE values, the higher measurement errors as compared to the reference instruments.

19


<sup>3</sup> Root Mean Square Error (RMSE): another metric to calculate measurement errors.

### Sensirion SEN54 vs South Coast AQMD Met Station (Temp; 5-min mean)




- The Sensirion SEN54 sensors showed very strong correlations with the corresponding South Coast AQMD Met Station data (0.97 < R<sup>2</sup> < 0.99)</li>
- Overall, the Sensirion SEN54 sensors overestimated the temperature measurement as recorded by South Coast AQMD Met Station
- The Sensirion SEN54 sensors seemed to track the diurnal temperature variations as recorded by South Coast AQMD Met Station

50



### Sensirion SEN54 vs South Coast AQMD Met Station (RH; 5-min mean)



## Discussion

- The three Sensirion SEN54 sensors' data recovery for all PM fractions was ~94.5%.
- The absolute intra-model variability was ~0.11, ~0.06 and ~0.05  $\mu$ g/m<sup>3</sup> for PM<sub>1.0</sub>, PM<sub>2.5</sub> and PM<sub>10</sub>, respectively.
- Reference instruments: strong correlations between FEM BAM and FEM T640 for PM<sub>2.5</sub> (R<sup>2</sup> ~0.75, 1-hr mean) and strong correlations between FEM BAM and T640 for PM<sub>10</sub> (R<sup>2</sup> ~0.85, 1-hr mean) mass concentration measurements
- The Sensirion SEN54 sensors showed strong correlations with the corresponding reference PM<sub>1.0</sub> data (0.88 < R<sup>2</sup> < 0.90, 1-hr mean), moderate to strong correlations with the corresponding reference PM<sub>2.5</sub> data (0.52 < R<sup>2</sup> < 0.79, 1-hr mean) and no to very weak correlations with the corresponding reference PM<sub>10</sub> data (0.09 < R<sup>2</sup> < 0.22, 1-hr mean). The sensors overestimated PM<sub>1.0</sub> and underestimated PM<sub>2.5</sub> and PM<sub>10</sub> mass concentrations as measured by the reference instruments.
- Temperature and relative humidity sensors showed very strong correlations with the South Coast AQMD Met Station T and RH data, respectively (R<sup>2</sup> ~ 0.98 for T and R<sup>2</sup> ~ 0.98 for RH) and overestimated the T and RH data as recorded by the South Coast AQMD Met Station.
- No sensor calibration was performed by South Coast AQMD staff for this evaluation.
- Laboratory chamber testing is necessary to fully evaluate the performance of these sensors under controlled T and RH conditions, and known target and interferent pollutants concentrations.

These results are still preliminary