SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT 21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS REPORT OF LABORATORY ANALYSIS

Page 1 of 2

TO	Jason Low	LABORATORY NO	1613330
	Atmospheric Measurements Manager		
	Science and Technology Advancement	DATE RECEIVED	05/13/2016
SAMI	PLE DESCRIBED AS		tel s
	24 hour sample	FACILITY ID NO	NA
	Canister 22472		
		REQUESTED BY	Sumner Wilson
SAMI	PLING LOCATION		
	Reseda Station	ST NO / PROJECT	NA
	18328 Gault St		
	18328 Gault St Los Angeles, CA 91335		

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	22472
Pressure (Torr)	740
CO, ppm	< 3
CH ₄ , ppm	2
CO ₂ , ppm	442
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 19/16

Approved By:

Soldmon Teffera, Acting Senior Manager

Laboratory Services 909-396-2199

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT 21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS

LABORATORY NO

1613330

24 hour sample Canister 22472

REQUESTED BY

Sumner Wilson

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

				Percent	QC Limit
CC106783	MDL	Theoretical	Measured	Difference	$\pm 5\%$ or ± 1
CO, ppmvC	0.3	1.92	2.34	21.88	PASS
CH ₄ , ppmvC	0.3	2.02	2.06	1.83	PASS
CO ₂ , ppmvC	0.3	1.57	2.42	54.20	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	2.07	2.08	PASS
NM/NEOC, ppmvC	0.4	2.03	2.08	2.34	PASS
				Percent	QC Limit
CC12628	MDL	Theoretical	Measured	Difference	$\pm 5\%$ or ± 1
CO, ppmvC	0.3	1036	1057	2.01	PASS
CH ₄ , ppmvC	0.3	1068	1048	-1.89	PASS
CO ₂ , ppmvC	0.3	1022	1042	1.93	PASS
C_2H_4 , ppmvC	0.4	NA	NA	NA	NA
C_2H_6 , ppmvC	0.4	1044	1046	0.15	PASS
NM/NEOC, ppmvC	0.4	1024	1022	-0.13	PASS

DATE ANALYZED REFERENCE NO:

05/13/2016 16QM2AB QM2-102

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT SAMPLE ANALYSIS REQUEST

\boxtimes	DIS
	INI
	LA
T A	RO

WC) #: 1	61333	0

ΓO: SCAQMD LAB: [OTHER:	: 🗆	16.46 (204)		
SOURCE NAME:	Southern Cali	fornia Gas Co.	I.D. No	0.	
Source Address: 128	01 Tampa Ave		City:	Porter Ran	ch
Mailing Address:		Ci	ty:	Zip:	91326
Analysis Requested by:	Sumner V	Wilson	Date:	5/13/16	5
Approved by:	Jason Low O	ffice:	1 3 3 3 1 -	Budget #:	44716
REASON REQUESTED Suspected Violation	Court/Hearing Board Rule(s)		_	Hazardous/Tox	ic Spill 🗌
Sample Collected by:	Qian Zhou	Date:	5/13/16	Time:	10:15pm
	palment of the last of		PAMS analysis		
	n Can#	Stort day		Stant was	E 1
City/Location	d Can#	Start day /	time/ duration	Start vac	End Press
Reseda Station	200 20000000000000000000000000000000000		0:00 / 24 hours	<-30"	
			<u> </u>		Press
			<u> </u>		Press
			<u> </u>		Press
	n 22472	5/12/16 / 0	0:00 / 24 hours	<-30"	Press +13
		5/12/16 / 0	<u> </u>		Press

Reseda Station - 18328 Gault St, Los Angeles, CA 91335

GPS (34.199225, -118.532743)